§ 866.3210

subpart E of part 807 of this chapter subject to the limitations in §866.9.

 $[47\ FR\ 50823,\ Nov.\ 9,\ 1982,\ as\ amended\ at\ 54\ FR\ 25046,\ June\ 12,\ 1989;\ 66\ FR\ 38791,\ July\ 25,\ 2001]$

§866.3210 Endotoxin assay.

- (a) *Identification*. An endotoxin assay is a device that uses serological techniques in whole blood. The device is intended for use in conjunction with other laboratory findings and clinical assessment of the patient to aid in the risk assessment of critically ill patients for progression to severe sepsis.
- (b) Classification. Class II (special controls). The special control for this device is the FDA guidance entitled "Class II Special Controls Guidance Document: Endotoxin Assay." See §866.1(e) for the availability of this guidance document.

[68 FR 62008, Oct. 31, 2003. Redesignated at 70 FR 53069, Sept. 7, 2005]

§866.3220 Entamoeba histolytica serological reagents.

- Identification. Entamoeba histolytica serological reagents are devices that consist of antigens and antisera used in serological tests to identify antibodies to Entamoeba histolytica in serum. Additionally, some of these reagents consist of antisera conjugated with a fluorescent dye (immunofluorescent reagents) used to identify Entamoeba histolytica directly from clinical specimens. The identification aids in the diagnosis of amebiasis caused by the microscopic protozoan parasite Entamoeba histolytica and provides epidemiological information on diseases caused by this parasite. The parasite may invade the skin, liver, intestines, lungs, and diaphragm, causing disease conditions such as indolent ulcers, an amebic hepatitis, amebic dysentery, and pulmonary lesions.
- (b) Classification. Class II (special controls). The device is exempt from the premarket notification procedures in subpart E of part 807 of this chapter subject to §866.9.

[47 FR 50823, Nov. 9, 1982; 47 FR 56846, Dec. 21, 1982, as amended at 63 FR 59226, Nov. 3, 1998]

§866.3225 Enterovirus nucleic acid assay.

- (a) Identification. An enterovirus nucleic acid assay is a device that consists of primers, probes, enzymes, and controls for the amplification and detection of enterovirus ribonucleic acid (RNA) in cerebrospinal fluid (CSF) from individuals who have signs and symptoms consistent with meningitis or meningoencephalitis. The detection of enterovirus RNA, in conjunction with other laboratory tests, aids in the clinical laboratory diagnosis of viral meningitis caused by enterovirus.
- (b) Classification. Class II (special controls). The special control is FDA's guidance document entitled "Class II Special Controls Guidance Document: Nucleic Acid Amplification Assay for the Detection of Enterovirus RNA." See §866.1(e) for the availability of this guidance document.

[74 FR 8, Jan. 2, 2009]

§866.3235 Epstein-Barr virus serological reagents.

- (a) Identification. Epstein-Barr virus serological reagents are devices that consist of antigens and antisera used in serological tests to identify antibodies to Epstein-Barr virus in serum. The identification aids in the diagnosis of Epstein-Barr virus infections and provides epidemiological information on diseases caused by these viruses. Epstein-Barr viruses are thought to cause infectious mononucleosis and have been associated with Burkitt's lymphoma (a tumor of the jaw in African children and young adults) and postnasal carcinoma (cancer).
- (b) Classification. Class I (general controls).

§ 866.3240 Equine encephalomyelitis virus serological reagents.

(a) Identification. Equine encephalomyelitis virus serological reagents are devices that consist of antigens and antisera used in serological tests to identify antobodies to equine encephalomyelitis virus in serum. The identification aids in the diagnosis of diseases caused by equine encephalomyelitis viruses and provides epidemiological information on these

Food and Drug Administration, HHS

viruses. Equine encephalomyelitis viruses are transmitted to humans by the bite of insects, such as mosquitos and ticks, and may cause encephalitis (inflammation of the brain), rash, acute arthritis, or hepatitis.

(b) Classification. Class I (general controls). The device is exempt from the premarket notification procedures in subpart E of part 807 of this chapter subject to §866.9.

[47 FR 50823, Nov. 9, 1982, as amended at 65 FR 2311, Jan. 14, 2000]

§ 866.3250 Erysipelothrix rhusiopathiae serological reagents.

Identification. *Erysipelothrix* rhusiopathiae serological reagents are devices that consist of antigens and antisera used in serological tests to identify Erysipelothrix rhusiopathiae from cultured isolates derived from clinical specimens. The identification aids in the diagnosis of disease caused by this bacterium belonging to the genus Erysipelothrix. This organism is responsible for a variety of inflammations of the skin following skin abrasions from contact with fish, shellfish, or poultry.

(b) Classification. Class I (general controls). The device is exempt from the premarket notification procedures in subpart E of part 807 of this chapter subject to the limitations in §866.9.

 $[47\ FR\ 50823,\ Nov.\ 9,\ 1982,\ as\ amended\ at\ 54\ FR\ 25046,\ June\ 12,\ 1989;\ 66\ FR\ 38791,\ July\ 25,\ 2001]$

§ 866.3255 Escherichia coli serological reagents.

(a) Identification. Escherichia coli serological reagents are devices that consist of antigens and antisera used in serological tests to identify Escherichia coli from cultured isolates derived from clinical specimens. Additionally, some of these reagents consist of Escherichia coli antisera conjugated with a fluorescent dye used to identify Escherichia coli directly from clinical specimens or cultured isolates derived from clinical specimens. The identification aids in the diagnosis of diseases caused by this bacterium belonging to the genus Escherichia, and provides epidemiological information on diseases caused by this microorganism. Although Escherichia coli constitutes the greater part of the

microorganisms found in the intestinal tract in humans and is usually non-pathogenic, those strains which are pathogenic may cause urinary tract infections or epidemic diarrheal disease, especially in children.

(b) Classification. Class I (general controls). The device is exempt from the premarket notification procedures in subpart E of part 807 of this chapter subject to the limitations in §866.9.

[47 FR 50823, Nov. 9, 1982, as amended at 54 FR 25046, June 12, 1989; 66 FR 38791, July 25, 2001]

§866.3270 Flavobacterium spp. serological reagents.

(a) Identification. Flavobacterium spp. serological reagents are devices that consist of antigens and antisera used in serological tests to identify Flavobacteriuim spp. from cultured isolates derived from clinical specimens. The identification aids in the diagnosis of disease caused by bacteria belonging to the genus Flavobacterium and provides epidemiological information on diseases caused by these microorganisms. Most members of this genus are found in soil and water and, under certain conditions, may become pathohumans. Flavobacterium genic to meningosepticum is highly virulent for the newborn, in whom it may cause epidemics of septicemia (blood poisoning) and meningitis (inflammation of the membranes of the brain) and is usually attributable to contaminated hospital equipment.

(b) Classification. Class I (general controls). The device is exempt from the premarket notification procedures in subpart E of part 807 of this chapter subject to the limitations in §866.9.

[47 FR 50823, Nov. 9, 1982, as amended at 54 FR 25046, June 12, 1989; 66 FR 38792, July 25, 2001]

§866.3280 Francisella tularensis serological reagents.

(a) Identification. Francisella tularensis serological reagents are devices that consist of antigens and antisera used in serological tests to identify antibodies to Francisella tularensis in serum or to identify Francisella tularensis in cultured isolates derived from clinical specimens. Additionally, some of these reagents consist of antisera conjugated